


## AS/NZS 4308:2008 - Certificate of Compliance

**ProScreen Cup – Manufactured by US Diagnostics Inc. USA**  
**Part No: PSCupA-6MBAU**

The ProScreen Cup was tested in accordance with Appendix B – “Verification of performance of on-site devices around the cut-off”, of the Australian and New Zealand Standard AS/NZS 4308:2008.

The ProScreen Cup was shown to satisfy the specified performance criteria when results are interpreted as described in the package insert (42141-ProII Revision 1 June, 2007).

Results of the verification procedure used to assess compliance of the ProScreen Cup have been documented and are available to any requesting authority and any accrediting agency.



Dr. John H. Vine  
Laboratory Director

11 November 2008