


## **AS/NZS 4308:2008 - Certificate of Compliance**

### **ProScreen Dip Card – Manufactured by US Diagnostics Inc. USA Part No: PSDA-6MBAU**

The ProScreen Dip Card was tested in accordance with Appendix B – “Verification of performance of on-site devices around the cut-off”, of the Australian and New Zealand Standard AS/NZS 4308:2008.

The ProScreen Dip Card was shown to satisfy the specified performance criteria when results are interpreted as described in the package insert (42131-ProII Revision 1 November, 2008).

Results of the verification procedure used to assess compliance of the ProScreen Dip Card have been documented and are available to any requesting authority and any accrediting agency.



Dr. John H. Vine  
Laboratory Director

14 July 2010